

บันทึกการตรวจสอบเครื่อง Defibrillator งานผู้ป่วยใน โรงพยาบาลหนองมะโมง

วัน/เดือน/ปี	Operational Check Report	ลงชื่อผู้ตรวจสอบ
16 มกราคม 2566	Efficia DFM100 866199 S/N: CN32609190 SW Rev: 2.00.21 American English Current Operational Check: 16 Jan 2023 08:29 Pass Last Operational Check: 12 Jan 2023 02:48 Pass	Current Test Results/ General System Test: Therapy Knob: Pass Charge Button: Pass Shock Button: Pass Audio Test: Pass Sync Button: Pass Therapy Delivery Test: Pass/Pads Leads ECG Test: Pass/ECG Cable Pads/Paddles ECG Test: Battery Test: SpO2 Test: Printer Test:
20 ม.ค 66	Current Operational Check: 29 Jan 2023 11:03 Pass Last Operational Check: 16 Jan 2023 08:29 Pass	Shock Button: Pass Audio Test: Pass Sync Button: Pass Therapy Delivery Test: Pass/Pads Leads ECG Test: Pass/ECG Cable Printer Test:
28 ก.ย 66.	Current Operational Check: 28 Feb 2023 13:42 Pass Last Operational Check: 28 Feb 2023 13:28 Fail/DX	Audio Test: Pass Sync Button: Pass Therapy Delivery Test: Pass/Pads Leads ECG Test: Pass/ECG Cable Printer Test: Pa
28 ก.ย 66. 22.004.	Current Operational Check: 28 Feb 2023 22:00 Pass Last Operational Check: 28 Feb 2023 21:59 Fail/DX	Shock Button: Pass Audio Test: Pass Sync Button: Pass Therapy Delivery Test: Pass/Pads Leads ECG Test: Pass/ECG Cable Printer Test:
16 มี.ค 66. 10:43 น.	SW Rev: 2.00.21 Current Operational Check: 16 Mar 2023 10:43 Pass Last Operational Check: 16 Mar 2023 10:42 Fail/DX	Audio Test: Pass Sync Button: Pass Therapy Delivery Test: Pass/Pads Leads ECG Test: Pass/ECG Cable

บันทึกการตรวจสอบเครื่อง Defibrillator งานไม้ไผ่ใน โรงพยาบาลนครพนม

Operational Check Report

Efficia DFM100 866199
 S/N: CN32609190
 SW Rev: 2.00.21 English (US)

Current Operational Check:
 28 May 2023 09:14 Pass

Last Operational Check:
 16 Mar 2023 10:43 Pass

Current Test Results:

General System Test: Pass
 Therapy Knob: Pass
 Charge Button: Pass
 Shock Button: Pass
 Audio Test: Pass
 Sync Button: Pass
 Therapy Delivery Test: Pass/Pads
 Leads ECG Test: Pass/ECG Cable

Pads/Paddles ECG Test: Pa:
 Battery Test: Pa:
 SpO2 Test: Pa:
 Printer Test: Pa:

วัน/เดือน/ปี

28/5/66

28/5/66	Operational Check Report Efficia DFM100 866199 S/N: CN32609190 SW Rev: 2.00.21 English (US) Current Operational Check: 28 May 2023 09:14 Pass Last Operational Check: 16 Mar 2023 10:43 Pass	Current Test Results: General System Test: Pass Therapy Knob: Pass Charge Button: Pass Shock Button: Pass Audio Test: Pass Sync Button: Pass Therapy Delivery Test: Pass/Pads Leads ECG Test: Pass/ECG Cable Pads/Paddles ECG Test: Pa: Battery Test: Pa: SpO2 Test: Pa: Printer Test: Pa: