

การติดตั้งเครื่องมืออุปกรณ์ทางการแพทย์และสาธารณสุขต้องเป็นไปตามข้อกำหนดของผู้ผลิต และต้องได้รับการทดสอบและตรวจสอบเครื่องมือก่อนการตรวจรับ เพื่อตรวจสอบสมบูรณ์พร้อมในการทำงานของเครื่องและความสามารถในการเชื่อมต่อกับระบบสนับสนุนของโรงพยาบาลได้อย่างปลอดภัย

เครื่องวัดความดันโลหิตอัตโนมัติชนิดสอดแขน



มีการสอบเทียบเครื่องมือก่อนการตรวจรับ



AND
A&D Medical

A&D Company Limited
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

A&D Engineering, Inc.
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

A&D Instruments Limited
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

A&D AUSTRALASIA PTY LTD.
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

A&D INSTRUMENTS INDIA PRIVATE LIMITED
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

CERTIFICATE

CERTIFICATE OF VERIFICATION AND CALIBRATION

This is to certify

That item **Automatic Blood pressure Monitors**

Model: **AND-TM2657P** Serial No. **2A4232681**


Has been assembled ,verified and calibrated as well as tested
prior to despatch and performed in accordance to our specifications



CERTIFICATE BY :SERVICE ENGINEER

6/1/2023

DATE



AND
A&D Medical

A&D Company Limited
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

A&D Engineering, Inc.
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

A&D Instruments Limited
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

A&D AUSTRALASIA PTY LTD.
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

A&D INSTRUMENTS INDIA PRIVATE LIMITED
128/128-129, Klongkum Road, Bangkok 10110 Thailand
Tel: +662 2561 1111 Fax: +662 2561 1112
www.andmedical.com

BLOOD PRESSURE MONITOR - CALIBRATION REPORT

BRAND: AND MODEL: TM2657P DESCRIPTION: Automatic Blood pressure Monitors SERIAL NUMBER: 2A4232681	Code: OUT-6601-003 ISSUE DATE: 6/1/2566
--	--

Reference Standard Instrument				
Item	Instrument	Brand	Model	S/N
1	NIBP Simulator	Cortec	MS 200	19120700005

NIBP Simulator Check

Standard BP Simulator	Test1	Test2	Test3	Average	Acceptable Range	PASS	FAIL
SYSTOLIC (60 mmHg)	62	61	60	61.00	57 mmHg - 63 mmHg	/	
DIASTOLIC (30 mmHg)	29	30	30	29.67	27 mmHg - 33mmHg	/	
PULSE (80 bpm)	81	80	80	80.33	77 bpm- 83 bpm	/	


Standard BP Simulator	Test1	Test2	Test3	Average	Acceptable Range	PASS	FAIL
SYSTOLIC (120 mmHg)	120	121	122	121.00	117 mmHg - 123 mmHg	/	
DIASTOLIC (80 mmHg)	81	80	82	81.00	77 mmHg - 83mmHg	/	
PULSE (80 bpm)	81	81	81	81.00	77 bpm- 83 bpm	/	

Standard BP Simulator	Test1	Test2	Test3	Average	Acceptable Range	PASS	FAIL
SYSTOLIC (200 mmHg)	202	200	200	200.67	197 mmHg - 203 mmHg	/	
DIASTOLIC (150 mmHg)	150	151	150	150.33	147 mmHg - 153mmHg	/	
PULSE (80 bpm)	80	81	81	80.67	77 bpm- 83 bpm	/	

Static Test	Test1	Test2	Test3	Average	Acceptable Range	PASS	FAIL
30 mmHg	31	30	31	30.67	27 mmHg - 33 mmHg	/	
150 mmHg	150	151	150	150.33	147 mmHg - 153mmHg	/	
280 bpm	281	280	281	280.67	277 bpm - 283 bpm	/	

Comments: 699-029-5201/b

Status: Passed
 Service Required



CALIBRATE BY: SERVICE ENGINEER